



Acceptance protocol

SN:

Date:

Operator:

Reflectances of the gray patches, background and slanted edges

Patch	Reference reflectance	Measured reflectance	Deviation	Measured optical density
1 (white)	0,8000		0,0000	0,10
2	0,7488		0,0000	0,13
3	0,6975		0,0000	0,16
4	0,6463		0,0000	0,19
5	0,5951		0,0000	0,23
6	0,5438		0,0000	0,26
7	0,4926		0,0000	0,31
8	0,4414		0,0000	0,36
9	0,3901		0,0000	0,41
10	0,3389		0,0000	0,47
11	0,2877		0,0000	0,54
12	0,2364		0,0000	0,63
13	0,1852		0,0000	0,73
14	0,1340		0,0000	0,87
15	0,0827		0,0000	1,08
16 (black)	0,0315		0,0000	1,50
Background	0,1778		0,0000	0,75
Slanted edge low	0,0912		0,0000	1,04
	0,3020		0,0000	0,52
Slanted edge high	0,0501		0,0000	1,30
	0,5248		0,0000	0,28

*Measured with X-Rite eXact

https://www.image-engineering.de/content/products/charts/IE_reference_data_accuracy.pdf

TE276AP20250414